

## CERTIFICATE OF INSPECTION

### 検査成績書

Product name / 品名: Column stand / コラム スタンド

Code No./コードNo.: 178-029

Serial No./製造No.: **035933**

Date of inspection / 検査年月日: 2020-6-30

Temperature/温度: 20°C

Issue No./発行No.: 33071062

Date of issue/発行年月日: 2020-7-15

#### Inspection result/検査結果


Appearance/外観		合格
Flatness GP 平面度	Permissible value/許容値	8.0 $\mu\text{m}$
	Measured value/測定値	1.1 $\mu\text{m}$
Column loconotivity / コラム 摺動力		合格
Judgment/総合判定		Passed/合格

\* We hereby certify that this product has been inspected and found to conform to the applicable Mitutoyo Inspection Standards

\* この商品は、当社の検査規格に適合していることを保証致します。

Measured value/測定値

Unit/単位:  $\mu\text{m}$



Mitutoyo Measuring Inst. (Suzhou) Co., Ltd.



Inspector / 検印

*Handwritten signature*

15004ZRev1.0