
# INTERNATIONAL STANDARD



First edition 2015-09-15

# Metallic materials — Leeb hardness test —

Part 1: **Test method** 

Matériaux métalliques — Essai de dureté Leeb — Partie 1: Méthode d'essai



Reference number ISO 16859-1:2015(E)



#### © ISO 2015, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Ch. de Blandonnet 8 • CP 401 CH-1214 Vernier, Geneva, Switzerland Tel. +41 22 749 01 11 Fax +41 22 749 09 47 copyright@iso.org www.iso.org

Page

# Contents

Forew	ord	iv					
1	Scope	1					
2	Normative references						
3	Principle						
4	Symbols, abbreviated terms, and designations						
5	Testing instrument	3					
6	Test piece   6.1 Shape   6.2 Thickness and mass   6.3 Surface preparation	<b>3</b> 3					
7	Procedure	4					
8	Uncertainty of the results	6					
9	Test report	6					
10	Conversions to other hardness scales or tensile strength values	6					
Annex	A (normative) Tables of correction factors for use in tests not conducted in direction of gravity	7					
Annex	B (normative) Procedure for periodic checking of testing instrument by the user	11					
Annex	c (informative) Uncertainty of the measured Leeb hardness values	12					
Annex	x <b>D</b> (informative) <b>Leeb hardness testing instruments</b>						
Biblio	graphy	20					

### Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see <a href="https://www.iso.org/directives">www.iso.org/directives</a>).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see <a href="https://www.iso.org/patents">www.iso.org/patents</a>).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary Information

The committee responsible for this document is ISO/TC 164, *Mechanical testing of metals*, Subcommittee SC 3, *Hardness testing*.

ISO 16859 consists of the following parts, under the general title *Metallic materials* — *Leeb hardness test*:

- Part 1: Test method
- Part 2: Verification and calibration of the testing devices
- Part 3: Calibration of reference test blocks

# Metallic materials — Leeb hardness test —

# Part 1: **Test method**

#### 1 Scope

This part of ISO 16859 covers the determination of a dynamic hardness of metallic materials using seven different Leeb scales (HLD, HLS, HLE, HLDL, HLD+15, HLC, HLG).

#### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16859-2, Metallic materials — Leeb hardness test — Part 2: Verification and calibration of the testing devices

ISO 16859-3, Metallic materials — Leeb hardness test — Part 3: Calibration of reference test blocks

#### **3** Principle

When testing hardness according to Leeb, a moving impact body collides at normal incidence with a surface and rebounds. The velocity of the impact body is measured before ( $v_A$ ) and after impact ( $v_R$ ). The energy amount absorbed by the test piece respectively dissipated in the test measures the dynamic Leeb hardness of the test piece. It is assumed that the impact body does not permanently deform.

The ratio of the impact and rebound velocity values gives the coefficient of restitution for the impact configuration and energy used. This coefficient represents the proportion of initial kinetic energy returned to the impact body within the contact time of the impact.

The hardness number according to Leeb, HL, is calculated as given in Formula (1)

$$HL = \frac{v_R}{v_A} \cdot 1\ 000$$

where

- *v*<sub>R</sub> is rebound velocity;
- *v*<sup>A</sup> is impact velocity.

By definition, the Leeb hardness is a ratio and thus becomes a quantity without dimensions.

#### 4 Symbols, abbreviated terms, and designations

**4.1** For most common Leeb scale and type of impact devices, see <u>Table 1</u>.

NOTE Other parameter values can be used based on the specific agreement between the parties.

(1)

Ch - l	II	it Designation	Parameters of types of impact devices						
Symbol	Unit		Da	S	E	DL	D+15	С	G
E <sub>A</sub>	mJ	Kinetic impact energy <sup>b</sup>	11,5	11,4	11,5	11,95	11,2	3,0	90,0
VA	m/s	Impact velocity	2,05	2,05	2,05	1,82	1,7	1,4	3,0
VR	m/s	Rebound velocity	0,615 - 1,824 5	0,82 - 1,886	0,615 - 1,886	1,1092 - 1,729	0,561 - 1,513	0,49 - 1,344	0,9 - 2,25
	mm	Maximum distance of ball indenter from test piece surface at velocity measurement	2,00	2,00	2,00	2,00	2,00	2,00	3,00
М	g	Mass of impact body incl. ball indenter	5,45	5,40	5,45	7,25	7,75	3,1	20,0
R	mm	Spherical radius of indenter ball	1,5	1,5	1,5	1,39	1,5	1,5	2,5
		Material of indenter	WC-Coc	Cd	PCDe	WC-Coc	WC-Co <sup>c</sup>	WC-Co <sup>c</sup>	WC-Coc
HL		Leeb hardness	HLD	HLS	HLE	HLDL	HLD+15	HLC	HLG
		Field of application	300 HLD - 890 HLD	400 HLS - 920 HLS	300 HLE - 920 HLE	560 HLDL - 950 HLDL	330 HLD+15 - 890 HLD+15	350 HLC - 960 HLC	300 HLG - 750 HLG

#### Table 1 — Symbols, dimensions, designations, and parameters of Leeb scales according to type of impact devices

b Impact vertically down, in direction of gravity, rounded.

с Tungsten-carbide cobalt.

d Ceramics.

Polycrystalline diamond.

The Leeb hardness number is followed by the symbol "HL" with one or more subsequent characters 4.2 representing the type of impact device.

#### EXAMPLE 570 HLD

Leeb hardness, HL, is measured using impact device type D in direction of gravity. Measurements using a different impact device type will deliver a different hardness number, as the result from Formula (1) depends on the parameters of each impact device type.

For testing in other directions, the measured hardness number will be biased. In such cases, a correction shall be applied in accordance with Annex A. If the test is not conducted in direction of gravity, the testing direction and correction shall be recorded, and the adjusted hardness number shall be given as the Leeb hardness result.

#### 5 Testing instrument

**5.1** The instrument used for Leeb hardness testing consists of an impact device (for an example, see <u>Annex D</u>) and an electronic measuring and indicating unit to determine the impact and rebound velocity of the impact body.

**5.2** The impact body consists of a spherical indenter and the holder of the indenter, see <u>Table 1</u>.

**5.3** The support ring shall be mounted tightly to the bottom of the impact device. Except for impact device type DL, the support surface shall be designed to prevent movement of the impact device during the test.

The support ring should be checked regularly, as wear can affect the readings. Specifically, the bottom surface of the support ring should be visually inspected. Deposits and dirt should be removed.

**5.4** The instrument shall meet the requirements of ISO 16859-2.

#### 6 Test piece

#### 6.1 Shape

**6.1.1** Leeb hardness testing can be done on test pieces of diverse shapes as long as the impact velocity vector is normal to the local surface region to be tested, and the support ring is stably placed on the test piece surface.

**6.1.2** Test pieces with curved surfaces (concave or convex) can be tested providing that the radius of curvature at the test location is not less than 50 mm for the impact device type G, or 30 mm for other impact devices, respectively.

**6.1.3** In all other cases, special support rings shall be used for a stable seating of the instrument on the test surface.

#### 6.2 Thickness and mass

The stiffness of the test piece, which is often determined by the local thickness and the mass of the test piece, shall be considered when selecting the impact device to be employed (see <u>Table 2</u>).

NOTE 1 Failure to provide adequate support will produce incorrect test results.

NOTE 2 Test pieces of mass less than the minimum indicated mass or pieces of sufficient mass with sections less than the minimum thickness require rigid support and/or coupling to a solid supporting body. Coupling refers to a method where the test piece is firmly connected to a much heavier support without straining or stressing the test piece. For example, an adhesive film can be applied between the test piece surface and the heavy support. This combination presents a larger combined mass to resist the impinging impact body. The coupling method can be used after comparison of the results with an uncoupled reference test piece of sufficient mass and thickness.

Type of impact devices	Minimum mass (no rigid support)	Minimum mass (rigid support)	Minimum thickness (uncoupled)	Minimum thickness (coupled)
	kg	kg	mm	mm
D, DL, D+15, S, E	5	2	25	3
G	15	5	70	10
С	1,5	0,5	10	1

Table 2 — Mass and thickness requirements of test piece

NOTE 3 Special geometries of the test piece, e.g. thin slabs or tube surfaces, can require additional support of the test location to also permit testing where the thickness of the test piece can be smaller than the minimum thickness given in Table 2. For example on tubes, the support requirement can be expressed in terms of the ratio of the tube diameter, *D*, to its wall thickness, *s*, (see References [2] to [4]), which is a measure of the sample stiffness. If no support can be applied, correction factors to the measured values can be determined in dependence of *D/s* (see Reference [4]).

#### 6.3 Surface preparation

The test surface shall be carefully prepared to avoid any alterations in hardness caused by heating during grinding or by work hardening during machining. It is recommended that the test surface be machined and polished to the surface finish as defined in Table 3. Any coatings, scale, contaminants, or other surface irregularities shall be completely removed. The surface shall be free from lubricants. The surface locations to be tested should not exceed the arithmetical mean roughness values,  $R_a$ , (also "centre line average") (see Reference [5]) given in Table 3 for each impact device (see References [2] or [4]).

Type of impact device	Maximum arithmetical mean surface roughness
	R <sub>a</sub>
	μm
D, DL, D+15, S, E	2,0
G	7,0
С	0,4

Table 3 — Recommended surface finish  $R_a$ 

#### 7 Procedure

**7.1** The daily verification defined in <u>Annex B</u> shall be performed before the first test of each day for each scale used.

**7.2** In general, the test should be carried out at ambient temperature within the limits of 10 °C to 35 °C. However, because temperature variation can affect the results, users of the Leeb test can choose to control the temperature within a tighter range, such as 23 °C  $\pm$  5 °C.

NOTE The temperature of the test material and the temperature of the hardness testing instrument can affect the test results. The test temperature can adversely affect the hardness measurement.

**7.3** Magnetic fields at the test location can affect the results of Leeb tests and must be avoided. Leeb hardness tests can be found particularly susceptible to ambient electromagnetic fields in the frequency range of a few kHz.

**7.4** The test piece and impact device shall not be moved during a test. The supporting surface shall be clean and free from contaminants (scale, lubricants, dirt, etc.).